

DUO  
STAR





# mintaka

## INNOVATIONS







INNOVATORS OF PRECISION BAROGRAPHIC SYSTEMS

## MENU MAP







### MAIN SCREENS





### MENU - SHOW VALUES




## MENU - SHOW GRAPH



## MENU - GPS, WIFI, STARX



GPS Summary  
1/09/17 8:36:42 UTC-6  
Lat 16°21.5807N  
Long 86°26.2459W  
Speed 0.3 Knots  
Course 01  
Pressure 1027.6 mb  
Tend 2.1 Tchar 2



WiFi Status  
Creating Network  
MintakaStar-348311  
IP 192.168.4.1  
PW: \*\*\*\*\*  
OK

When in  
WxService Mode

WiFi Status  
Creating Network  
MintakaWx  
IP 192.168.4.1  
PW: \*\*\*\*\*  
OK


When in  
WxService Mode



Access WiFi Setup  
Enter Pin 0000  
OK Cancel



WiFi Status  
Creating Network  
MintakaWx  
IP 192.168.4.1  
PW: JWS01000  
Mode: No Change  
OK Cancel



StarX Report  
5/15/18 14:00:58 UTC-1  
24.8°C wb 18.2°C  
RH 52% dp 14.3°C  
1005.7 mb  
IP: 192.168.4.2

5/15/18 13:37:00 UTC-1  
StarX Report  
192.168.4.2  
Station Pressure hPa  
OK

4/16/18 10:37:47  
StarX has missed  
4 Reports  
OK



24.8 C 52 %  
1005.70 mb  
5/15/18 14:00:58 UTC-1

## MENU - SET-UP

Show Values  
 Show Graph  
 GPS, WiFi, StarX  
 Set-up

When in  
WxService Mode

Access Setup

Enter Pin 0000  
OK Cancel

Time  
 Units, Timezone  
 Pressure Settings  
 System

Time UTC-6  
 Jan 08 2017  
 10:29:29  
 OK Cancel

Time  
 Units, Timezone  
 Pressure Settings  
 System

Pressure: mb  
 Elevation: Feet  
 Time: 24hr UTC -7  
 OK Cancel

Time  
 Units, Timezone  
 Pressure Settings  
 System

Sea Level Pres.  
 Station Pres. Adj

Set Elevation Offset  
 00,000  
 OK Cancel

Set Pressure mb  
 1027.0  
 OK Cancel

Sea Level Pres.  
 Station Pres. Adj

Sensor: 1025.36  
 Offset by: 0.00  
 To read: 1025.00  
 OK Cancel

Time  
 Units, Timezone  
 Pressure Settings  
 System

About  
 Reset  
 Display, USB  
 Wx Service Mode

About

Mintaka Star  
 Version 3.0.7384  
 Serial #S84BFA24

Reset

Reset  
 Pressure history  
 OK Cancel

Display, USB

Backlite Timeout 1 m  
 Contrast 80  
 Baud Rate 57600  
 OK Cancel

Wx Service Mode

Enable WxService  
 Enter Pin 0000  
 OK Cancel

Entered  
 WxService Mode  
 Rebooting Now  
 OK

Disable WxService  
 Enter Pin 0000  
 OK Cancel

Exited  
 WxService Mode  
 Rebooting Now  
 OK